





# Glucose 2,000mg/L

## DATA SHEET

### Instructions for use

- 1. Immerse an indicator strip in the test solution for 2 seconds.
- 2. Shake off excess solution.
- 3. Wait 2 minutes.
- 4. Compare the colour produced on the strip against the colour chart.

#### **Measuring range**

0 - 100 - 250 - 500 - 1,000 - 2,000mg/dL

#### **Application**

Glucose test strips detect glucose levels in vegetables and fruit juices. They are also used in the educational sector for food science studies, urine analysis and osmosis/diffusion experiments.

#### **Shelf Life**

2 years from manufacturing date when stored between 2 -  $8^{\circ}$ C.

E.&O.E.

ABN 57 622 896 593

ACN 622 896 593

#### Contents

100 strips

#### **Dimensions**

Strip dimension: 83mm x 5mm

#### Notes

Keep any unused strips in the original container. Do not store where they are liable to be exposed to acid or alkaline gases or vapours. Store in a cool dry place.

E.&O.E.