



COBI REHAB[®]
The Positioning Company

XXL-Rehab Patientstol

XXL-Rehab Patient/Dinner chair

Anvendelse
Use

Max 325 kg

Brugsanvisning
User manual

DK

Bruges til:

- Bariatriske brugere på plejecentre, i hjemmet eller i venteværelser.

UK

For use by:

- Bariatric users at home, at care centers or in waiting rooms.



Brugsanvisning for XXL-Rehab Patientstol

ADVARSEL



Produktet må ikke anvendes uden grundig instruktion fra uddannet personale.

Bruges produktet til personer, der vejer mere end den maksimale brugervægt, er der risiko for at beskadige produktet.

Cobi Rehab påtager sig intet ansvar for eventuelle skader på produktet eller personskader forårsaget af forkert installation, forkert montering eller forkert anvendelse af produktet.

Kontakt Cobi Rehab med det samme, hvis der er dele, som er beskadigede eller mangler. Brug aldrig egne erstatningsreservedele.

Cobi Rehab kan ikke holdes ansvarlig for fejl eller ulykker, der opstår efter reparationer, der er lavet uden skriftlig tilladelse fra Cobi Rehab.

 Brugsanvisningen skal læses grundigt før XXL-Rehab Patientstol tages i brug. Ret specielt opmærksomheden mod Advarsler og Sikkerhedsvejledninger mærket med .


Desuden må stolen kun betjenes af uddannet personale eller af personer, der har fået instruktion i dens brug, da der ellers er fare for, at brugeren kan komme til skade.

Anvendelse

XXL-Rehab Patientstolen er en specialdesignet stol, der giver god siddekomfort til bariatriske brugere op til 325 kg.

Patientstolen bruges som spisebordsstol, patientstol i venteværelse, på sygehuse og på institutioner.

Bruges XXL-Rehab Patientstolen til personer, der vejer mere end 325 kg, kan stolen vælte eller gå i stykker.


 For at undgå at XXL-Rehab Patientstolen vælter, må den kun bruges på hårdt underlag med en maksimal hældning på 3°.

Før brug

Boltene til justering af ben og ryg skal monteres korrekt og strammes i hullerne på siden af stolen, før den tages i brug. Forkert montering og løse bolte kan betyde, at brugeren falder ud af stolen.

XXL-Rehab Patientstolen må kun monteres og justeres af specialuddannet personale eller af Cobi Rehab.

Under brug

 Vær altid opmærksom på usikre løft eller bevægelser, når der forflyttes til eller fra XXL-Rehab Patientstolen.

Sæderyggen kan indstilles i tre forskellige dybder. Justering af sædedybden påvirker ikke stolens stabilitet.

Vedligeholdelse og rengøring

Stolens stel, sæde og armlæn kan rengøres med vand og et almindeligt rengøringsmiddel.

Sædet kan også tages af, hvis der er behov for desinficering med fx sprit. Sørg for at fastgøre sædet korrekt igen efter rengøring.

Kontrollér regelmæssigt alle bolte og møtrikker. Spænd dem om nødvendigt.

⚠ Manglende kontrol af boltene som anvist kan forårsage funktionsfejl.

Garanti

Cobi Rehab garanterer, at XXL-Rehab Patientstol er fejlfri, når den leveres. Skulle der mod forventning være mangler eller skade på stolen ved modtagelsen, så kontakt Cobi Rehab med det samme.

Garantiperioden for XXL-Rehab Patientstol er et år fra købsdato. Fakturaen gælder som garantibevis.

Garantien gælder ikke:

- produkter, hvor serienummer, batchnummer eller lignende er blevet fjernet eller beskadiget væsentligt
- produkter, hvor der er foretaget reparation af uautoriseret personale
- normal vedligeholdelse fx rengøring, tilpasning, smøring og opdatering af udstyr eller dele.

Garantien gælder ikke, hvis Cobi Rehab vurderer, at XXL-Rehab Patientstol er blevet håndteret på en ureglementeret måde, fx hvis brugervægten overskrides.

Ved skade på grund af forkert brug, faktureres der både service og anvendte reservedele.

Garantien, beskrevet ovenover, er den eneste gældende garanti. Cobi Rehab kan ikke holdes ansvarlige for skader af nogen slags.

Specifikationer

Brugervægt max.	325 kg
Total længde	80-101 cm
Total bredde over armlæn	74 / 84 cm
Total højde	79-94 cm
Sædebredde	61 / 71 cm
Sædedybde justerbar	40-52 cm
Sædehøjde justerbar	44-59 cm
Sæderyg højde	37 cm
Afstand fra gulv til underside af sæde	33-48 cm
Afstand mellem armlæn	65 / 75 cm
Højde sæde til armlæn	28 cm
Vægt	16,5 / 17,5 kg
Ramme	Pulverlakeret krom molybdæn stål
Sæde/ryg/armlæn	Polyuretan
ISO - klassifikation	ISO 09 33 03
Produktets forventede levetid er	7 år

Bortskaffelse

Produkter, der ikke længere er i brug, skal bortskaffes på forsvarlig vis. Kontakt kommunen for vejledning om korrekt bortskaffelse.



XXL-Rehab Patientstol er produceret og testet i overensstemmelse med den internationale DS/EN 12182:2000 standard.



Scan koden og få flere produktinformationer.



PRODUCENT OG FORHANDLER

Cobi Rehab
Fuglebækvej 1D
DK-2770 Kastrup
Danmark
Tlf.: +45 70 25 25 22
cobirehab.com

09/2019/CP

Instructions for XXL-Rehab Patient/Dinner chair

CAUTION



The XXL-Rehab Patient/Dinner chair must not be used without proper instruction by a qualified professional.

Using the XXL-Rehab Patient/Dinner chair for individuals whose weight exceeds the maximum weight-bearing capacity may pose a risk for damaging the product.

Cobi Rehab assumes no responsibility for any damage to the product or personal injury caused by improper installation, assembly or use of this product.

If components are damaged or missing, contact Cobi Rehab immediately. Do not use substitute parts.


Cobi Rehab cannot be held liable for faults or accidents occurring after repairs by service staff made without written authorisation from Cobi Rehab.


 To ensure proper use of the XXL-Rehab Patient/Dinner chair, please read these instructions thoroughly before using the Patient/Dinner chair. In doing so, pay particular attention to the warnings and safety instructions marked with .

XXL-Rehab Patient/Dinner chair may only be assembled and adjusted by qualified staff who has been instructed in the proper use of the Patient/Dinner chair. Incorrect assembling and adjustment can injure the user.

Use

The XXL-Rehab Patient/Dinner chair is designed for bariatric users weighing up to 325 kg. The chair is used at hospitals, in waiting rooms or at home at the dinner table.


 If used by someone weighing more than 325 kg, there is a risk that the XXL-Rehab Patient/Dinner chair will break or sustain damage.

 To avoid tipping, always use the XXL-Rehab Patient/Dinner chair on level, hard flooring with a maximum slope of 3°.

Before use

The bolts for adjusting the legs and backrest must be properly mounted and tightened on the side of the XXL-Rehab Patient/Dinner chair.

During use

 Pay attention to unsafe lifts or movements during transfer from and to the XXL-Rehab Patient/Dinner chair.

The seat back can be adjusted in three different positions. Adjusting the depth setting on the seat back does not affect the stability of the XXL-Rehab Patient/Dinner chair.

The surface temperature of the XXL-Rehab Patient/Dinner chair can increase or decrease when exposed to external sources of heat or cold.

Maintenance and cleaning

The frame components and the seat and armrests may be cleaned with water and ordinary detergent. To disinfect the seat use alcohol. Make sure to fasten the seat properly again, if removed for disinfection.

Regularly check all bolts and nuts, and retighten if necessary.

 Failure to inspect the bolts may cause malfunction.

Warranty

When delivered, the XXL-Rehab Patient/Dinner chair is free from defects. Should, contrary to all inspections, there be any defects or damage to the Patient/Dinner chair, please contact Cobi Rehab immediately.

The XXL-Rehab Patient/Dinner chair is covered by a 1-year warranty from the date of purchase. The invoice serves as warranty certificate.

The warranty does not apply to:

- products whose serial number, batch number etc. have been removed or considerably damaged
- products that have been repaired by unauthorised staff
- normal wear and tear, for example cleaning, adjustment, lubrication and update of equipment or parts.

The warranty is void if Cobi Rehab assesses that XXL-Rehab Patient/Dinner chair has been used for purposes other than its intended use, for example, if its weight capacity is exceeded.

In case of incorrect use, both service and spare parts will be invoiced.

The warranty described above is the only valid warranty.

Cobi Rehab may not be held liable for damages of any kind.

Specifications

Maximum weight capacity is	325 kg
Total length	80-101 cm
Total outer width on armrests	74 / 84 cm
Total height	79-94 cm
Seat width	61 / 71 cm
Seat depth	40-52 cm
Seat height	44-59 cm
Height of backrest	37 cm
Distance from floor to seat underside	33-48 cm
Distance between armrests	65 / 75 cm
Seat to armrest	28 cm
Weight	16.6 / 17.6 kg
Frame	Powder-lacquered chromium molybdenum steel
Seat / back / armrests	Polyurethane
ISO - classification	ISO 09 33 03
Expected product service life	7 years

Disposal

Products that can no longer be used are to be disposed of in a sustainable way. Please contact the relevant authorities for information on proper disposal.



XXL-Rehab Patient/Dinner chair is manufactured and tested in accordance with the international DS/EN 12182:2000 standard.



MANUFACTURER AND DEALER

Cobi Rehab
Fuglebækvej 1D
DK-2770 Kastrup
Denmark
Tel.: +45 70 25 25 22
cobirehab.com

09/2019/CP



COBI REHAB[®]

The Positioning Company

COBI REHAB
Fuglebækvej 1D
DK-2770 Kastrup
Denmark
Tel. +45 7025 2522
cobirehab.com