

Pool Lab / Puresilk Free Chlorine Reagent #1 Colour Change

Free Chlorine Reagent #1 colour will alter over time.

Your new Free Chlorine #1 reagent may be clear or light magenta in colour, this is Okay and is expected.

If the colour is dark Magenta as per image below, the reagent has expired and may no longer provide accurate Free Chlorine readings.

As a general rule, if you cannot read the label printed on the bag, the reagent has expired.

To maximise the shelf life the reagents, store in a cool dark location.

