



S-7XTRA® Microbiocidal Activity – Medical Devices

AUST R 463988 | AUST L 232011 | AUST L 232014

S-7XTRA® has been rigorously tested for use in medical environments using industry-standard testing methodologies to simulate real-world use. Refer to the table below for targeted pathogens and required contact times.

S-7XTRA® is proven bactericidal, virucidal, fungicidal, yeasticidal, and sporicidal in dirty conditions and is intended for use on non-invasive medical devices (excluding 'FRESH') against pathogens in the formats indicated below.

| Pathogen | Contact Time | RTU | RTU | Wipes |
|---|--------------|--------------|---------------|-------|
| | | 5L & Squeeze | Trigger Spray | |
| Bacteria | | | | |
| Enterococcus hirae | 5 mins | ✓ | ✓ | ✓ |
| Escherichia coli (E. coli) | 5 mins | ✓ | ✓ | ✓ |
| Pseudomonas aeruginosa | 5 mins | ✓ | ✓ | ✓ |
| Staphylococcus aureus | 5 mins | ✓ | ✓ | ✓ |
| Methicillin-Resistant Staphylococcus aureus (MRSA) | 5 mins | ✓ | | ✓ |
| Proteus vulgaris | 5 mins | ✓ | | ✓ |
| Salmonella typhimurium | 5 mins | ✓ | | ✓ |
| Fungi & Yeast | | | | |
| Candida albicans | 5 mins | ✓ | ✓ | ✓ |
| Aspergillus brasiliensis | 20 mins | ✓ | | ✓ |
| Viruses | | | | |
| Vaccina Virus Ankara (MVA) | 5 mins | ✓ | ✓ | ✓ |
| SARS-CoV-2 (COVID-19) (Murine hepatitis virus surrogate) | 5 mins | ✓ | ✓ | ✓ |
| Poliovirus (Murine norovirus S99 surrogate) | 5 mins | ✓ | ✓ | ✓ |
| Adenovirus | 5 mins | ✓ | ✓ | ✓ |
| Norovirus (Murine norovirus S99 surrogate) | 5 mins | ✓ | ✓ | ✓ |
| Spores | | | | |
| Bacillus subtilis | 5 mins | ✓ | ✓ | |



NOTES:

- a) ✓ indicates the format has been proven effective and is intended for use against the target pathogen.
- b) RTU 5L and squeeze includes use of S-7XTRA® by flooding, mopping, or pouring.
- c) S-7XTRA® Concentrate must be diluted in accordance with the instructions for *“Intermediate-high risk environments including non-invasive medical devices”* in 02-05-05 IFU – Directions for Use – S-7XTRA® Concentrate to achieve the results indicated.